



<sup>1</sup>H NMR

Triethyl phosphite

[122-52-1]



Solvent: CDCl<sub>3</sub>  
Concentration: 16.7 mg/ 1.0 ml  
Reference substance: TMS  
Sample temperature: 30.0 °C  
Resonance frequency: 400.136 MHz  
Spectral width: 4629.6 Hz  
Data points (FID; spec.): 64 k; 64 k

Flip angle; Pulse width: 45 °; 6.2 μs  
No. of scans; Rep. time: 32; 23.1 s  
Weighting; Line broad.: exp.; 0.1 Hz  
Spectral resolution: 0.3 Hz (TMS)  
Instrument: Bruker AMX-400  
Source reference: 921112-116  
Signature: *Markku Mesilaakso*  
Markku Mesilaakso

nuc.	δ [ppm]	J [Hz]
a	3.876	8.0 (P); 7.1 (b)
b	1.269	0.5 (P)

5-3-157

OPCW Code:05-3-0157(p1/1)