



¹H NMR

Diethylphosphite

[762-04-9]



Solvent: CDCl₃
Concentration: 20 µl/ 0.8 ml
Reference substance: TMS
Sample temperature: 30.0 °C
Resonance frequency: 400.132 MHz
Spectral width: 4629.6 Hz
Data points (FID; spec.): 64 k; 64 k

Flip angle; Pulse width: 45 °; 6.1 µs
No. of scans; Rep. time: 32; 23.1 s
Weighting; Line broad.: exp.; 0.1 Hz
Spectral resolution: 0.3 Hz (TMS)
Instrument: Bruker AMX-400
Source reference: 971010-4128
Signature: *Markku Mesilaakso*
Markku Mesilaakso

nuc.	δ [ppm]	J [Hz]
a	4.148	-10.2 (a'); 7.1 (b); 8.9 (P)
a'	4.157	7.1 (b); 9.3 (P)
b	1.368	0.3 (P)
c	6.811	692.3 (P)

Iterative analysis

OPCW Code:05-3-0161(p1/1)

5-3-161